

Survival rates and causes of revision in cemented primary total knee replacement

A REPORT FROM THE NORWEGIAN ARTHROPLASTY REGISTER
1994–2009

Ø. Gøthesen, B. Espehaug, L. Havelin, G. Petursson, S. Lygre, P. Ellison, G. Hallan, O. Furnes

This is an unedited, pre-publication version of the paper. Reproduced with permission and copyright © of the British Editorial Society of Bone and Joint Surgery [Gøthesen Ø, Espehaug B, Havelin L, et al. Survival rates and causes of revision in cemented primary total knee replacement: a report from the Norwegian Arthroplasty Register 1994-2009. J Bone Joint Surg [Br] 2013;95-B:636-642.]

Abstract

We evaluated the rates of survival and cause of revision of seven different brands of cemented primary total knee replacement (TKR) in the Norwegian Arthroplasty Register during the years 1994 to 2009. Revision for any cause, including resurfacing of the patella, was the primary endpoint. Specific causes of revision were secondary outcomes. Three posterior cruciate-retaining (PCR) fixed modular-bearing TKRs, two fixed non-modular bearing PCR TKRs and two mobile-bearing posterior cruciate-sacrificing TKRs were investigated in a total of 17 782 primary TKRs. The median follow-up for the implants ranged from 1.8 to 6.9 years. Kaplan-Meier 10-year survival ranged from 89.5% to 95.3%. Cox's relative risk (RR) was calculated relative to the fixed modular-bearing Profix knee (the most frequently used TKR in Norway), and ranged from 1.1 to 2.6. The risk of revision for aseptic tibial loosening was higher in the mobile-bearing LCS Classic (RR 6.8 (95% confidence interval (CI) 3.8 to 12.1)), the LCS Complete (RR 7.7 (95% CI 4.1 to 14.4)), the fixed modular-bearing Duracon (RR 4.5 (95% CI 1.8 to 11.1)) and the fixed non-modular bearing AGC Universal TKR (RR 2.5 (95% CI 1.3 to 5.1)), compared with the Profix. These implants (except AGC Universal) also had an increased risk of revision for femoral loosening (RR 2.3 (95% CI 1.1 to 4.8), RR 3.7 (95% CI 1.6 to 8.9), and RR 3.4 (95% CI 1.1 to 11.0), respectively). These results suggest that aseptic loosening is related to design in TKR.

The aim of this study was to investigate the rate of survival and causes of revision for seven brands of cemented primary total knee replacement (TKR) registered in the Norwegian Arthroplasty Register (NAR) between 1994 and 2009. The brands are the currently and historically among the most commonly used both in Norway and around the world^{1, 2}. The study was limited to cemented implants without patellar resurfacing, and the data reflect the results of the average surgeon. We accept that pooling of data from many surgeons, with different experience, patient volumes and skills, may give good external validity but may also hide the effect of a learning curve and any positive effect that may be related to high volumes undertaken by some surgeons.

We also investigated whether survival was brand specific or related to particular types of design.

Patients and Methods

Data from patients registered in the NAR during this time were evaluated. The registration of hip replacements in the NAR started in 1987 and was expanded to include TKRs and the replacement of other joints in 1994^{3,4}. The completeness of the registration was estimated by Espehaug et al⁵ to be 99% of all primary TKRs and 97% of all revision procedures between 1999 and 2002. Any complete or partial removal/exchange of the implant, or insertion of a component (including a patellar component), was considered a revision procedure. The unique identification number of all Norwegian residents facilitates linking the revisions to the primary operations.

All TKRs were cemented and were inserted without patellar components. Differences between the designs were predominantly on the tibial side; two were mobile-bearing TKRs (LCS Classic and LCS Complete (DePuy, Warsaw, Indiana), both rotating platform), two were non-modular fixed bearing TKRs (AGC Universal and AGC Anatomic; both Biomet, Warsaw, Indiana), and three were modular fixed-bearing TKRs (Duracon; Stryker, Portage, Michigan; NexGen; Zimmer, Warsaw, Indiana; and Profix; Smith & Nephew, Memphis, Tennessee). The mobile-bearing TKRs were posterior cruciate ligament (PCL) sacrificing, and the others were PCL retaining.

Implant designs not in use after 2004, and those which were used in < 500 cases, were excluded (Fig. 1). TKRs introduced with computer-navigation were excluded because the technique was not widely used for the TKRs which were selected. Posterior-stabilised implants were excluded because of relatively low numbers (the Profix Conforming Plus was regarded as posterior stabilised). The inclusion criteria were met by 2118 AGC Universal, 1190 AGC Anatomic, 1090 Duracon, 778 NexGen, 6276 Profix, 2606 LCS Classic and 3714 LCS Complete TKRs.

Statistical analysis

Revision for any cause was the primary endpoint. Specific causes for revision and types of revision were secondary outcomes. Descriptive analyses were used to assess the baseline characteristics of the various brands (Table I). Information on deaths or emigrations up to 31 December 2009 was retrieved from the National Population Register. The survival times of unrevised TKRs were taken at the last date of observation (date of death or emigration, or 31 December 2009). Median follow-up was calculated with the reverse Kaplan–Meier method⁶. Unadjusted survival curves for the various brands were constructed using the Kaplan–Meier method, and stopped when < 50 knees remained at risk. Survival percentages after five and ten years' follow-up are reported. Cox's multiple regression model was used to calculate hazard rate ratios (RR), adjusted for potential confounding by age, gender, pre-operative diagnosis (osteoarthritis or other diagnoses) and previous knee surgery (yes/no). The RR estimates are presented with 95% confidence intervals (CI) and p-values reported relative to the Profix TKR, which was the most common TKR in Norway in the last decade. A sub-analysis was performed to present the risk estimates of the category of design relative to fixed modular-bearing designs. We tested the proportional hazards assumption of the Cox model based on scaled Schoenfeld residuals.^{7,8} With revision for any reason as the endpoint, the assumption was found valid for the factors 'prosthesis brand' with the Profix implant as the reference brand ($p \geq 0.1$) and 'design category' with fixed modular bearing as the reference category ($p \geq 0.6$). Bilateral TKRs were included in the study. Although this might imply a violation of the assumption of independent observations in the survival analyses, studies have shown that the impact on statistical precision is minor for both hip⁹ and knee replacements.¹⁰

PASW Statistics version 18 (IBM SPSS, IBM Corporation, Armonk, New York) and R v2.13.0 R ((The R Foundation for Statistical Computing, <http://www.R-project.org> 2008) were used for the statistical analyses, and $p < 0.05$ was considered statistically significant.

The NAR has approval from the Norwegian Data Inspectorate to collect patient data on condition of based on a concession and the a written consent of the patient.

Results

The study groups did not differ markedly with respect to age, gender, laterality or diagnosis (Table I). The median follow-up ranged from 1.8 to 6.9 years depending on the implant (Table II). The Cox's regression analyses and the Kaplan-Meier curves showed that the Duracon, LCS Classic, LCS Complete and AGC Universal brands had a higher risk of revision (RR 1.3 to 2.6) and a statistically significantly lower survival (89.5% to 94.0%) than the Profix TKR (95.3%) (Table II, Fig. 2). The NexGen and the AGC Anatomic TKR performed in a similar manner to the Profix. A sub-analysis of TKRs performed in the latest time period, after 2004, showed a higher risk of revision for the two mobile-bearing implants (RR 1.3 (95% CI 1.0 to 1.7)), but not for monobloc implants (RR 1.0 (95% CI 0.7 to 1.4)) compared with the fixed-bearing implants.

There was an increased risk of revision for aseptic tibial loosening in the LCS Classic, LCS Complete and the Duracon TKRs compared with the Profix (RR 6.8 (95% CI 3.8 to 12.1), RR 7.7 (95% CI 4.1 to 14.4) and RR 4.5 (95% CI 1.8 to 11.1), respectively) (Table III and Fig. 3a). These implants also had an increased risk of revision for aseptic femoral loosening (RR 2.3 (95% CI 1.1 to 4.8), RR 3.7 (95% CI 1.6 to 8.9) and RR 3.4 (95% CI 1.1 to 11.0), respectively) (Fig. 3b). Also, the AGC Universal TKR had an increased risk of revision for aseptic tibial loosening (RR 2.5 (95% CI 1.3 to 5.1)) compared with the Profix. The risk of revision due to deep infection was higher for all TKRs except the LCS Classic, compared with the Profix (RR from 1.8 to 3.7). The risk of revision due to polyethylene wear and to malalignment was higher in the Duracon TKRs (RR 16.6 (95% CI 4.9 to 56.7) and RR 8.7 (95% CI 3.7 to 20.4), respectively). However, the number of revisions for these reasons was low ($n = 10$ and $n = 10$, respectively). The LCS Classic had a higher risk of revision due to dislocation of the polyethylene (RR 3.7 (95% CI 1.2 to 11.1)). The AGC Universal had a higher risk of revision due to pain (RR 2.1 (95% CI 1.5 to 3.0)) and dislocation of the patella (RR 8.0 (95% CI 1.6 to 39.6)), whereas the LCS Complete and LCS Classic had a lower risk of revision due to pain as the only cause of revision (RR 0.4 (95% CI 0.2 to 0.8)). Insertion of a patellar component was the most frequent revision operation performed for pain.

Using the fixed modular-bearing category as the reference, for the three categories of design we found an increased risk of revision due to aseptic loosening of the tibial tray in the mobile-bearing (RR 4.8 (95% CI 3.2 to 7.3)) and the monobloc category (RR 1.9 (95% CI 1.1 to 3.3)). Aseptic loosening of the femoral component was more common in the mobile-bearing category (RR 2.5 (95% CI 1.4 to 4.4)). Further, we included only the most used subtypes of the implants in the analysis, but the results did not change.

In order to minimise the effect of a learning curve, we performed a sub-analysis that only included TKRs from hospitals having inserted > 100 . The risk of revision for aseptic tibial loosening was still

higher for the LCS Classic (RR 5.8 (95% CI 3.3 to 10.2) and the LCS Complete (RR 6.8 (95% CI 3.6 to 12.8)) compared with the Profix TKR. The type of cement did not influence survival. The mean operating time ranged from 85 minutes for the AGC Anatomic to 105 minutes for the NexGen TKR (Table I).

In order to preclude any time dependency, we analysed the one-year and five-year Kaplan-Meier overall survival rates and the Cox's regression hazard rate ratios. The differences in survival of the various brands did not change markedly over time.

Discussion

The Duracon, LCS Classic, LCS Complete and AGC Universal brands had lower survival than the Profix, whereas the NexGen and AGC Anatomic TKRs did not. Increased risk of revision for aseptic loosening of the tibial and femoral components was the major reason for the inferior performance. The AGC Universal was more likely to be revised because of pain than the other brands, and LCS Complete and LCS Classic were less likely to be revised for this reason. The risk of revision for deep infection was higher for all brands, except the LCS Classic, than for the Profix.

The implants with a higher risk of aseptic loosening represent different design principles, so no common thread was apparent. For example, the fixed non-modular bearing AGC Universal was inferior to the Profix, but the AGC Anatomic was not.

Revision because of pain was rare with mobile-bearing implants, which is consistent with the theory that rotation of the mobile bearing improves patellar tracking.¹¹ The AGC Anatomic, with right/left femoral components, has replaced the AGC Universal, and its good results are consistent with data from the Australian Arthroplasty Registry showing similar revision rates for monoblock and fixed-bearing TKRs after ten years.²

The risk of revision due to dislocation of the polyethylene bearing was higher for the LCS Classic than for the Profix, but not for the LCS Complete. In our study most of the mobile-bearing LCS TKRs sacrificed the PCL, whereas the fixed modular and fixed non-modular TKRs were PCL retaining (Table I).

This study focused on the causes of revision and found the highest risk of revision to be in the LCS TKRs, for both aseptic tibial and femoral loosening. Other studies have shown good survival and clinical results of mobile-bearing designs,¹¹⁻¹⁴ but these studies did not compare mobile with fixed bearings. The inferior results of the mobile-bearing TKRs in our study are consistent with data from the Australian Joint Replacement Registry² and from the Southern California Permanente Medical Group.^{15,16}

The aim of the mobile-bearing design was to combine low constraint forces with low contact stresses, theoretically reducing polyethylene wear and aseptic loosening.¹⁷ Early fixed-bearing designs had unsatisfactory function and range of movement, and it was claimed that the biomechanics of the mobile-bearing design were closer to those of a normal knee, and would improve function and longevity.¹⁸ Dislocation of the polyethylene was a problem in the early years of the mobile-bearing TKR, but as the technique and instruments evolved, this complication became rare.¹⁹ However, there is no strong evidence that any mobile-bearing design is superior to a fixed bearing with regard to pain, function, range of movement or failure rate.^{20,21} It is claimed that wear

of the polyethylene in the modular fixed bearing and the mobile bearing at the tibial interface may lead to peri-prosthetic osteolysis and loosening.^{22, 23} This so-called backside wear is eliminated in the fixed non-modular (monobloc) design, but the modularity option is lost. The monobloc design has excellent survival in several studies,²⁴⁻²⁷ but most surgeons prefer the modular fixed bearing.

A retrieval study evaluating 48 mobile bearings concluded that wear was as severe as that in fixed modular-bearing designs.²³ Similar polyethylene wear was found for a mobile-bearing rotating platform and a fixed modular bearing in an in vitro study.²⁸ Another in vitro study, however, concluded that the wear rate of the fixed bearing was four times higher than for the rotating platform,²⁹ but in two meta-analyses no differences in the incidence of radiolucent lines or clinical outcome were found.^{21, 30} Recent reports from the NAR did not show differences in pain, function or survival for the LCS Classic, or survival for the LCS Complete, compared to the AGC Universal TKR.^{31, 32} Differences in geometry and undersurface texture in the two mobile-bearing TKRs might explain why they differ in outcome.³³ All the mobile-bearing TKRs in this study were 'no keel' subtypes, and there might have been less resistance to rotational forces with this design compared to those with a keel (Table IV). The higher risk of revision for aseptic loosening of the tibial and femoral components in the LCS Classic and LCS Complete must be further investigated, focusing on wear and shear forces at the prosthesis–cement–bone interfaces.

The inferior results reported here for the Duracon TKR differ from those reported from the Australian Arthroplasty Register.² A possible explanation could be that in 2005 the Duracon TKR was introduced in one geographical region of Norway as a result of a tender process, and therefore the local surgeons were obliged to go through a learning process.

In conclusion, differences in the causes of revision were brand specific. The assumption that fixed modular-bearing implants are more at risk of loosening due to polyethylene wear than mobile-bearing designs was not supported by this study.

Supplementary material

A table detailing the use of subtypes of implants within each brand and two Kaplan-Meier survival curves showing cumulative survival at i) one and ii) five years.

Acknowledgements

The authors would like to thank Paul Johan Høl (PhD) for valuable input regarding interpretation of the results.

Funding

The corresponding author received funding from the Research Council of Norway.

Declaration of interest

No benefits in any form have been received or will be received from a commercial party related directly or indirectly to the subject of this article.




Figure 1. Selection chart showing inclusions and exclusions of cases. There may be more than one exclusion criteria per case (* rare combinations of implants: Profix mobile-bearing (n = 12), AGC Dual (52), various combinations of LCS (n = 565)).




Figure 2. Kaplan-Meier survival analysis of the various brands with revision for any reason as the endpoint.




Figure 3a




Figure 3b

Figure 3a and 3b show Kaplan-Meier analyses of the various brands with a) tibial and b) femoral loosening as the endpoint.

Table I. Demographic data (ACL – anterior cruciate ligament, PCL – posterior cruciate ligament)

Characteristic	Implant							
	AGC Anatomic	AGC Universal	Duracon	LCS Complete	LCS Classic	NexGen	Profix	
Patients (n)	1190	2118	1090	3714	2606	778	6276	
Male (n, %)	450 (37.8)	650 (30.7)	352 (32.3)	1219 (32.8)	718 (27.6)	273 (35.1)	1950 (31.1)	
Right knee (n, %)	637 (53.5)	1149 (54.2)	596(54.7)	2004 (54.0)	1437 (55.1)	394 (50.6)	3422 (54.5)	
Mean (sd) age (yrs)	69.7 (9.1)	71.0 (9.2)	70.7 (9.3)	69.6 (9.6)	71.5 (9.0)	69.2 (10.5)	70.0 (10.0)	
Age group (n, %)								
≤ 60 years	169 (14.2)	259 (12.2)	141 (12.9)	597 (16.1)	299 (11.5)	130 (16.7)	1015 (16.2)	
61 to 70 years	393 (33.0)	575 (27.1)	330 (30.3)	1214 (32.7)	706(27.1)	235(30.2)	1846 (29.4)	
71 to 80 years	492 (41.3)	962(45.4)	446 (40.9)	1394(37.5)	1174(45.0)	316(40.6)	2434 (38.8)	
> 80 years	136(11.4)	322 (15.2)	173 (15.9)	509 (13.7)	427 (16.4)	97(12.5)	981(15.6)	
Diagnosis (n, %)								
Primary osteoarthritis	1062(89.5)	1832 (86.9)	950 (87.5)	3338 (90.1)	2268 (87.4)	674 (86.7)	5325(85.2)	
Other	124(10.5)	276(13.1)	136(12.5)	366(9.9)	328(12.6)	103(13.3)	928(14.8)	

Hospitals using this design (n)	11	29	18	35	30	19	40
Mean (sd) operation time (min)	85 (23)	96 (23)	98 (30)	97 (24)	101 (35)	105 (54)	92 (32)
Cement with antibiotics (n, %)	1187(99.7)	2115(99.9)	1089(99.9)	3711(99.9)	2592(99.5)	778 (100)	6200(98.8)
Patients with previous operations on the knee (n, %)	311(26.2)	456(21.6)	333 (30.6)	1128 (30.4)	688 (26.4)	205(26.4)	1542(24.6)
Intact ACL pre-operatively (n, %)	904 (76.0)	1664(78.8)	792(73.3)	2995 (80.7)	1940 (74.6)	542(69.7)	5246(83.6)
Intact PCL post-operatively (n, %)	1027(91.9)	1950 (96.3)	911(87.7)	250(6.9)	353(13.9)	649(95.2)	5941 (97.0)

Table II. Kaplan-Meier survival by implant brand of cemented primary total knee replacements without patellar resurfacing, reported to the Norwegian Arthroplasty Register between 1994 and 2009, with revision for all causes as the endpoint (CI, confidence interval)

Implant	Total (n)	Revised (n, %)	Median follow-up (yrs)	5 years		10 years		Relative risk (95% CI)*	p-value
				Survival (%; 95% CI)	At risk (n)	Survival (%; 95% CI)	At risk (n)		
Profix	6276	195 (3.1)	4.5	96.3 (95.7 to 96.9)	2575	95.3 (94.5 to 96.1)	51	1	
Duracon	1090	56 (5.1)	1.8	93.3 (91.1 to 95.5)	247	89.5 (86.1 to 92.9)	117	2.6 (1.9 to 3.4)	< 0.001
NexGen	778	25 (3.2)	3.2	94.7 (92.5 to 96.9)	159	-**	4	1.2 (0.8 to 1.9)	0.3
LCS Classic	2606	129 (5.0)	6.6	95.6 (94.8 to 96.4)	1898	94.0 (92.8 to 95.2)	261	1.3 (1.0 to 1.6)	0.017
LCS Complete	3714	102 (2.7)	1.9	94.9 (93.5 to 96.3)	61	-***	0	1.5 (1.1 to 1.9)	0.002
AGC Universal	2118	121 (5.7)	6.9	94.7 (93.7 to 95.7)	1436	92.6 (91.2 to 94.0)	369	1.6 (1.3 to 2.0)	< 0.001
AGC Anatomic	1190	29 (2.4)	2.7	96.5 (95.1 to 97.9)	119	-****	0	1.1 (0.7 to 1.6)	0.7

* Cox regression with adjustment for age, gender, diagnosis and previous surgery. ** last revision at 4.65 years, *** last revision at 4.31 years, **** last revision at 3.89 years.

Table III. Causes of revision by incidence and Cox's relative risk (RR) for cemented total knee replacements without patellar resurfacing reported to the Norwegian Arthroplasty Register between 1994 and 2009. There may be more than one cause of revision reported in each case. Statistically significant differences compared with the Prefix implant, are marked with bold (C), confidence interval)

		Implant									
Cause of revision	Prefix	Duracon	NexGen	LCS Complete	LCS Classic	AGC Universal	AGC Anatomic				
Incidence (n, %)											
Aseptic loosening (femur)	12 (0.2)	4 (0.4)	1 (0.1)	10 (0.3)	16 (0.6)	5 (0.2)	1 (0.1)				
Aseptic loosening (tibia)	15 (0.2)	7 (0.6)	4 (0.5)	32 (0.9)	58 (2.2)	17 (0.8)	3 (0.3)				
Dislocation (patella)	2 (0.0)	5 (0.5)	-	-	-	6 (0.3)	1 (0.1)				
Dislocation (other)	5 (0.1)	-	-	1 (0.0)	9 (0.3)	1 (0.0)	-				
Instability	31 (0.5)	11 (1.0)	5 (0.6)	10 (0.3)	13 (0.5)	15 (0.7)	5 (0.4)				
Malalignment	12 (0.2)	10 (0.9)	2 (0.3)	5 (0.1)	13 (0.5)	4 (0.2)	2 (0.2)				
Deep infection*	31 (0.5)	15 (1.4)	11 (1.4)	33 (0.9)	21 (0.8)	22 (1.0)	11 (0.9)				
Fracture affecting implant	8 (0.1)	1 (0.1)	-	4 (0.1)	8 (0.3)	-	-				
Pain**	68 (1.1)	3 (0.3)	3 (0.4)	11 (0.3)	14 (0.5)	51 (2.4)	7 (0.6)				
Polyethylene wear	4 (0.1)	10 (0.9)	-	3 (0.1)	1 (0.0)	2 (0.1)	-				
Stiffness	12 (0.2)	5 (0.5)	-	5 (0.1)	1 (0.0)	1 (0.0)	-				
Other	27 (0.4)	10 (0.9)	1 (0.1)	11 (0.3)	7 (0.3)	7 (0.3)	-				

RR (95% CI)											
Aseptic loosening (femur)	1	3.4 (1.1 to 11.0)	0.9 (0.1 to 6.8)	3.7 (1.6 to 8.9)	2.3 (1.1 to 4.8)	0.8 (0.3 to 2.4)	0.8 (0.1 to 6.3)				
Aseptic loosening (tibia)	1	4.5 (1.8 to 11.1)	2.9 (0.9 to 8.6)	7.7 (4.1 to 14.4)	6.8 (3.8 to 12.1)	2.5 (1.3 to 5.1)	1.7 (0.5 to 5.9)				
Dislocation (patella)	1	19.3 (3.7 to 100.3)	0	0	0	8.0 (1.6 to 39.6)	3.1 (0.3 to 34.7)				
Dislocation (other)	1	0	0	0.5 (0.1 to 4.0)	3.7 (1.2 to 11.1)	0.5 (0.1 to 4.6)	0				
Instability	1	3.5 (1.7 to 7.0)	1.5 (0.6 to 3.8)	1.0 (0.5 to 2.1)	0.8 (0.4 to 1.6)	1.2 (0.7 to 2.3)	1.3 (0.5 to 3.3)				
Malalignment	1	8.7 (3.7 to 20.4)	1.8 (0.4 to 7.9)	1.4 (0.5 to 4.1)	2.1 (0.9 to 4.6)	0.9 (0.3 to 2.7)	1.4 (0.3 to 6.3)				
Deep infection*	1	3.7 (2.0 to 6.9)	3.3 (1.6 to 6.5)	2.6 (1.6 to 4.3)	1.4 (0.8 to 2.5)	1.8 (1.1 to 3.2)	2.4 (1.2 to 4.7)				
Fracture affecting implant	1	0.8 (0.1 to 6.6)	0	1.2 (0.4 to 4.1)	1.9 (0.7 to 5.1)	0	0				
Pain†	1	0.4 (0.1 to 1.4)	0.4 (0.1 to 1.3)	0.4 (0.2 to 0.8)	0.4 (0.2 to 0.8)	2.1 (1.5 to 3.0)	0.7 (0.3 to 1.5)				
Polyethylene wear	1	16.6 (4.9 to 56.7)	0	4.0 (0.8 to 19.9)	0.3 (0.0 to 3.0)	0.8 (0.1 to 4.3)	0				
Stiffness/Other	1	3.7 (1.8 to 7.7)	0.3 (0.0 to 2.4)	1.1 (0.5 to 2.2)	0.6 (0.3 to 1.3)	0.8 (0.3 to 1.8)	0				

* deep infection rules out aseptic loosening

** pain as the only cause of revision


Reference List

1. **Engesaeter LB, Furnes O, Havelin LI, Fenstad AM** Annual report from the Norwegian Arthroplasty Register 2010. http://nrlweb.ihelse.net/eng/Report_2010.pdf (date last accessed 5 March 2013). In: 2013.
2. **Graves S, Davidson D, Tomkins A** Australian Orthopaedic Association National Joint Replacement Registry: annual report 2011. <http://www.dmac.adelaide.edu.au/aoanjrr> (date last accessed 15 January 2013). In: 2013.
3. **Furnes O, Espehaug B, Lie SA, Vollset SE, Engesaeter LB, Havelin LI** Early failures among 7,174 primary total knee replacements: a follow-up study from the Norwegian Arthroplasty Register 1994-2000. *Acta Orthop Scand* 2002;73:117-29.
4. **Havelin LI, Engesaeter LB, Espehaug B, Furnes O, Lie SA, Vollset SE** The Norwegian Arthroplasty Register: 11 years and 73,000 arthroplasties. *Acta Orthop Scand* 2000;71:337-53.
5. **Espehaug B, Furnes O, Havelin LI, Engesaeter LB, Vollset SE, Kindseth O** Registration completeness in the Norwegian Arthroplasty Register. *Acta Orthop* 2006;77:49-56.
6. **Schemper M, Smith TL** A note on quantifying follow-up in studies of failure time. *Control Clin Trials* 1996;17:343-6.
7. **Grambsch PM** Goodness-of-fit and diagnostics for proportional hazards regression models. *Cancer Treat Res* 1995;75:95-112.
8. **Ranstam J, Karrholm J, Pulkkinen P, Makela K, Espehaug B, Pedersen AB, Mehnert F, Furnes O** Statistical analysis of arthroplasty data. *Acta Orthop* 2011;82:258-67.
9. **Lie SA, Engesaeter LB, Havelin LI, Gjessing HK, Vollset SE** Dependency issues in survival analyses of 55,782 primary hip replacements from 47,355 patients. *Stat Med* 2004;23:3227-40.
10. **Robertsson O, Ranstam J** No bias of ignored bilaterality when analysing the revision risk of knee prostheses: analysis of a population based sample of 44,590 patients with 55,298 knee prostheses from the national Swedish Knee Arthroplasty Register. *BMC Musculoskelet Disord* 2003;4:1.
11. **Callaghan JJ, Squire MW, Goetz DD, Sullivan PM, Johnston RC** Cemented rotating-platform total knee replacement. A nine to twelve-year follow-up study. *J Bone Joint Surg Am* 2000;82:705-11.
12. **Callaghan JJ, Wells CW, Liu SS, Goetz DD, Johnston RC** Cemented rotating-platform total knee replacement: a concise follow-up, at a minimum of twenty years, of a previous report. *J Bone Joint Surg Am* 2010;92:1635-9.

13. **Kim YH, Kim JS, Park JW, Joo JH** Comparison of the low contact stress and press fit condylar rotating-platform mobile-bearing prostheses in total knee arthroplasty: a prospective randomized study. *J Bone Joint Surg Am* 2011;93:1001-7.
14. **Sorrells RB, Voorhorst PE, Murphy JA, Bauschka MP, Greenwald AS** Uncemented rotating-platform total knee replacement: a five to twelve-year follow-up study. *J Bone Joint Surg Am* 2004;86-A:2156-62.
15. **Namba RS, Inacio MC, Paxton EW, Ake CF, Wang C, Gross TP, Marinac-Dabic D, Sedrakyan A** Risk of revision for fixed versus mobile-bearing primary total knee replacements. *J Bone Joint Surg Am* 2012;94:1929-35.
16. **Paxton EW, Namba RS, Maletis GB, Khatod M, Yue EJ, Davies M, Low RB, Jr., Wyatt RW, Inacio MC, Funahashi TT** A prospective study of 80,000 total joint and 5000 anterior cruciate ligament reconstruction procedures in a community-based registry in the United States. *J Bone Joint Surg Am* 2010;92 Suppl 2:117-32.
17. **Buechel FF, Pappas MJ** The New Jersey Low-Contact-Stress Knee Replacement System: biomechanical rationale and review of the first 123 cemented cases. *Arch Orthop Trauma Surg* 1986;105:197-204.
18. **O'Connor JJ, Goodfellow JW** Theory and practice of meniscal knee replacement: designing against wear. *Proc Inst Mech Eng H* 1996;210:217-22.
19. **Buechel FF, Sr.** Long-term followup after mobile-bearing total knee replacement. *Clin Orthop Relat Res* 2002:40-50.
20. **Kim YH, Kook HK, Kim JS** Comparison of fixed-bearing and mobile-bearing total knee arthroplasties. *Clin Orthop Relat Res* 2001:101-15.
21. **Oh KJ, Pandher DS, Lee SH, Sung Joon SDJ, Lee ST** Meta-analysis comparing outcomes of fixed-bearing and mobile-bearing prostheses in total knee arthroplasty. *J Arthroplasty* 2009;24:873-84.
22. **Ingram JH, Stone M, Fisher J, Ingham E** The influence of molecular weight, crosslinking and counterface roughness on TNF-alpha production by macrophages in response to ultra high molecular weight polyethylene particles. *Biomaterials* 2004;25:3511-22.
23. **Kelly NH, Fu RH, Wright TM, Padgett DE** Wear damage in mobile-bearing TKA is as severe as that in fixed-bearing TKA. *Clin Orthop Relat Res* 2011;469:123-30.
24. **Gill GS, Joshi AB** Long-term results of cemented, posterior cruciate ligament-retaining total knee arthroplasty in osteoarthritis. *Am J Knee Surg* 2001;14:209-14.
25. **Himanen AK, Belt E, Nevalainen J, Hamalainen M, Lehto MU** Survival of the AGC total knee arthroplasty is similar for arthrosis and rheumatoid arthritis. Finnish Arthroplasty Register report on 8,467 operations carried out between 1985 and 1999. *Acta Orthop* 2005;76:85-8.
26. **Ritter MA, Meneghini RM** Twenty-year survivorship of cementless anatomic graduated component total knee arthroplasty. *J Arthroplasty* 2010;25:507-13.

27. **Worland RL, Johnson GV, Alemparte J, Jessup DE, Keenan J, Norambuena N** Ten to fourteen year survival and functional analysis of the AGC total knee replacement system. *Knee* 2002;9:133-7.
28. **Haider H, Garvin K** Rotating platform versus fixed-bearing total knees: an in vitro study of wear. *Clin Orthop Relat Res* 2008;466:2677-85.
29. **Delport HP, Sloten JV, Bellemans J** Comparative gravimetric wear analysis in mobile versus fixed-bearing posterior stabilized total knee prostheses. *Acta Orthop Belg* 2010;76:367-73.
30. **Smith TO, Ejtehadi F, Nichols R, Davies L, Donell ST, Hing CB** Clinical and radiological outcomes of fixed- versus mobile-bearing total knee replacement: a meta-analysis. *Knee Surg Sports Traumatol Arthrosc* 2010;18:325-40.
31. **Lygre SH, Espehaug B, Havelin LI, Vollset SE, Furnes O** Does patella resurfacing really matter? Pain and function in 972 patients after primary total knee arthroplasty. *Acta Orthop* 2010;81:99-107.
32. **Lygre SH, Espehaug B, Havelin LI, Vollset SE, Furnes O** Failure of total knee arthroplasty with or without patella resurfacing. *Acta Orthop* 2011;82:282-92.
33. **Crossett L** Evolution of the low contact stress (LCS) complete knee system. *Orthopedics* 2006;29:S17-S22.


Supplementary material



Cox’s multiple regression hazard rate ratios the first year

Implant	RR (95% CI)	p-value
Profix	1	
Duracon	2.8 (1.7 to 4.5)	< 0.001
LCS Complete	1.6 (1.1 to 2.3)	0.021

Figure i



Cox's multiple regression hazard rate ratios the first five year

Implant	RR (95% CI)	p-value
Profix	1	
AGC Universal	1.7 (1.2 to 2.0)	< 0.001
Duracon	2.3 (1.7 to 3.2)	< 0.001
LCS Complete	1.4 (1.1 to 1.8)	0.003
LCS Classic	1.3 (1.0 to 1.6)	0.042

Figure ii

Figure i and ii. Kaplan-Meier curves showing survivorship by brand at i) one year and ii) five years, hazard rates significantly different from Profix.

Table i. Subtypes registered for each prosthesis brand and number revised

Subtype	Tibia	Femur	Revised
NexGen			
Precoat PMMA stemmed	379		8
Precoat PMMA pegged	5		1
Option	391		16
Precoat CR		31	2
Option CR		594	22
Option CRA		4	0
CR Flex Option		140	1
CR Flex gender-specific		6	0
LCS			
PCR porocoat	1		0
PCR textured	11		5
Rotating platform porocoat	7		0
Rotating platform textured	2539		128
Rev rotating platform porocoat	45		2
Mod rev	3		0
Porocoat		121	5
Textured		2474	128
Rev porocoat		3	0
Mod rev		8	0
AGC			
V2 interlok	3252		147
Interlok	61		2
Anatomic porous		2	0
Anatomic interlok		1188	29
Universal interlok		57	3
V2 interlok		2061	118
LCS Complete			
No keel MBT	3676		99
With keel MBT	1		0
MBT revision	20		1
Small, standard, large (+)		3609	94
Mod/revision-unconstrained s,m,st,l (+)		5	0
Profix			
Non-porous	6276	6276	195
Duracon			
Porous with screw fixation	4		0
Porous with stem	3		0
Porous/resurf	2		0
Cruciform/porous	110		13
Univ/porous	2		0
Univ/non-porous	9		2
Cruciform/non-porous	904		38
Bead, PCA	14		0
Resurf, PCA	36		3
Cruciform, beaded	1		0
Porous		166	14
Non-textured		780	36
Porous Modular		12	0
Non-textured Modular		121	5
Monolithic(PS)		3	0